



## **DECLARATION OF CONFORMITY**

In accordance with Law 268/2014 Comp., On medical devices, in conjunction with Government Decree No. 54/2015 Comp., On technical requirements for medical devices, in accordance with Law No. 22/1997, Comp., on technical requirements for products and Council Directive 93/42 / EEC on medical devices.

Manufacturer information:

## **GOOD MASK s.r.o.**

Balbínova 529/1, Praha 2 - Vinohrady 120 00, Czech Republic

## **Product identification data:**

Reference: Surgical Mask OM1

Intended use: The intended purpose of the mask is to protect the wearer's environment by filtering exhaled air and reducing the transmission of infection between people.

## **Medical device class: I** (non-sterile and measurement functions)

The manufacturer declares that the properties of the aforementioned medical device meet all the requirements established by Law No. 268/2014 Comp., Government Decree No. 54/2015 Comp. and Directive 93/42 / EEC, and that the medical device is safe, effective and suitable for the provision of healthcare for its intended use. The manufacturer further declares that he has taken measures to ensure the conformity of the medical device marketed with the essential requirements and the manufacturer's technical documentation.

The conformity assessment was carried out according to the following standards:

ČSN EN 15223-1 ČSN EN 13795 + A1 ČSN EN 14683

In Prague 1. 7. 2020

Mgr. Martin Ladyr, CEO

GOOD MASK s.r.o. Balbínova 1/529 120 00 Praha 2 1ČO: 09112286 www.goodmask.cz